



THE UNIVERSITY OF  
SOUTHERN MISSISSIPPI  
1848 • 1948 • 2018

# THE UNIVERSITY OF SOUTHERN MISSISSIPPI

January 5<sup>th</sup>, 2022

## ADDENDUM 1 TO BID 22-25

This addendum provides answers to questions submitted by prospective bidders and adds a coversheet to the bid packet on the website. The answers to the questions below are shown in red.

---


Steve Ballew, Director of Procurement and Contract Services

1. In the last bullet, it mentions “control and data acquisition software.” Does this just mean ability to remotely control the unit and use its inherent ability to use scripting to automate analysis functions? Wanted to ensure no software development was expected here outside of a native ability. **Native ability is what we are looking for here.**
2. Bullet #2 states “receiver step attenuator on one port.” Just wanted to confirm it is ONLY needed on one port and not potentially on an additional second port. **Provide quote with step attenuators on two ports.**
3. Bullet #4 in the sub-section states “A third and fourth internal source would save wear and tear of antenna switching and increases testing throughput.” Is this something that is required with the initial delivered unit, or just an option which could be added at a later time? Both can be done but would affect quoting and cost. **Add this to the quote**
4. A revised copy of the bid packet titled “Bid 22-25” has been placed on the website. Please follow this link to view the revised bid packet titled “Bid 22-25”: <https://www.usm.edu/procurement-contract-services/current-bids-and-sole-source-notices.php> Please use the coversheet that is now in the revised packet.

## PROCUREMENT SERVICES

118 College Drive #5003 • Hattiesburg, MS • 39406-0001

*Hattiesburg • Long Beach • Ocean Springs • Biloxi • John C. Stennis Space Center*



# **THE UNIVERSITY OF SOUTHERN MISSISSIPPI**

## **PROCUREMENT SERVICES**

118 College Drive #5003 • Hattiesburg, MS • 39406-0001

*Hattiesburg • Long Beach • Ocean Springs • Biloxi • John C. Stennis Space Center*