



12 November 2024

Addendum 2 for RFP 25-18 Respirometer

This addendum provides answers to questions submitted by prospective bidders.
The University's answers are shown in **RED**.

Amber Floyd

Buyer, Procurement, and Contract Services

Could you kindly provide more details about the number of chambers the end user requires?

We would like 24 channels. 12 of the channels require a shaking apparatus appropriate for ASTM D6691.